



# EC-CERTIFICATE

(Full quality assurance system)

## DQS Medizinprodukte GmbH

hereby certifies that the company

## TRIGO care International GmbH

In den Wiesen 3  
51674 Wiehl  
Germany

has implemented and maintains a full quality assurance system which applies to the products at every stage from design to final controls.

An audit, documented in a report, performed by DQS, has verified that this quality assurance system fulfils the requirements of

### Annex II – excluding Section 4 of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

sterile Medical devices ( class IIb ) for wound care:  
Hydrogel dressings, Alginate dressings, Foam dressings, Hydrocolloid dressings, Lipocolloid dressings, silver containing wound dressings

The manufacturer is subject to surveillance according to Annex II, Section 5. The CE marking with the Notified Body Number (0297) may be affixed on the devices listed in the certificate. An EC Design Examination Certificate according to Annex II, Section 4 is required for class III devices covered by this certificate. In case of class Is devices the certificate is restricted to the aspects of manufacture concerned with securing and maintaining sterile conditions. In case of class Im devices the certificate is restricted to the aspects of manufacture concerned with the conformity of the products with metrological requirements.

Certificate registration No.	470698 MR2
Certificate unique ID	170528511
Effective date	2011-08-05
Expiry date	2015-10-09
Frankfurt am Main	2011-08-05

Frank Graichen  
Managing Director

Stefan Hofmann  
Head of Certification Body

August-Schanz-Straße 21, 60433 Frankfurt am Main, Tel. +49 (0) 69 95427-263, [medical.devices@dqs.de](mailto:medical.devices@dqs.de)

**DQS Medizinprodukte GmbH is a Notified Body according to Council Directive 93/42/EEC concerning medical devices with the Notified Body Number 0297.**